

Kraków, 26 January 2012

Ref. No.: 120126-JMP-4-ZO

**Request for Quotation  
No.  
120126-JMP-4-ZO**

*Regarding:* The supply of laboratory equipment for the LifeScience Park in Krakow at ul. Bobrzyńskiego 14

**Subject of the Contract**

The subject of the contract is the supply of laboratory equipment as specified below.

1. **Detailed description of laboratory equipment:**  
**Franz diffusion cells with stirring and temperature control systems**

ITEM NO.	QUANTITY	PRODUCT NAME	PRODUCT DESCRIPTION
1.	20	Jacketed Franz diffusion cells	A clear glass Franz cell, 9 mm in diameter. The cells should come with a thermostatic water jacket with the receptor volume of at least 4 ml, stir bar and pinch clamp (e.g. the 4G-G1000905 device manufactured by SES Analytical Systems or equivalent).
2.	2	Magnetic holder/stirrer	The set of cells (20 pieces) should include 2 compact magnetic stirrers. The stirrer must have the capacity to evenly stir up to 9 vessels. It must enable temperature and heating control. The device should offer the acryl holders and tubings for connections to cell jackets (e.g. the PG-V9-CA device manufactured by SES Analytical Systems or equivalent).

strona 1 / 3



## 2. Additional terms:

- a) Delivery date not later than 9 weeks from the conclusion of the contract.
- b) The parameters specified for individual items represent the minimum parameter thresholds, the parameters of the products offered may not fall below those thresholds.
- c) The Contracting Authority reserves the right to award a supplementary contract.
- d) The Bidder may submit a bid for either one part of the contract or for the complete contract. A partial bid shall cover the entirety of an item listed above in the "ITEM NO." column.
- e) The Contracting Authority reserves the right to purchase part of the components listed.
- f) No advance shall be paid. Payment for the supplied devices shall be made within 21 days from the receipt of a VAT invoice issued within 7 days from the date the product delivery and acceptance report is signed, confirming the delivery of equipment to the Contracting Authority.
- g) If the terms of guarantee require the start-up and calibration of the equipment by qualified technical service staff, the cost of such start-up should be included in the price of the device.

## 3. Submission of bids

Bids shall be submitted at the secretary's office of the Jagiellonian Centre for Innovation (at ul. Bobrzyńskiego 14, 30-348 Kraków, Poland) by 10 February 2012, 11:00 CET. The request for quotation shall be deemed received at the date it is published on the Contracting Authority's website.

A declaration acknowledging the Bidder's awareness of the procurement regulations available at the Jagiellonian Centre for Innovation's website at [www.jci.pl](http://www.jci.pl) shall be enclosed with the bid.

## 4. Bid evaluation criteria

The bids shall be evaluated in terms of their compliance with the request for quotation. The winning bid shall be the one that offers the lowest price for a given part of the contract.

Having evaluated the bids, the Contracting Authority shall notify the Bidders immediately of the results of the evaluation procedure.

Jagiellońskie Centrum Innowacji Sp. z o.o.  
Prezys Zarządu  
*Paweł Błachno*  
Paweł Błachno

strona 2 / 3



.....  
city/town, date

### DECLARATION

I, the undersigned, ....., holder of ID No. ...., acting on behalf of ....., declare that I have read the "Procurement regulations of Jagiellonian Centre for Innovation regarding the project of Extension of Jagiellonian LifeScience Park and Technology Incubator as part of Measure 5.3 of the Innovative Economy Operational Programme" of 25 March 2011, adopted by a resolution No. 1 of the Board of 25 March 2011.

-----

Contractor's signature

strona 3 / 3



UNIA EUROPEJSKA  
EUROPEJSKI FUNDUSZ  
ROZWOJU REGIONALNEGO

**LIFE SCIENCE PARK**  
Bobrzyńskiego 14, 30-348 Kraków

T: +48 12 297 46 00  
F: +48 12 297 46 46

E: [jci@jci.pl](mailto:jci@jci.pl)  
W: [www.jci.pl](http://www.jci.pl)

